

## **CE-Declaration of Conformity**

For the following  
equipment:

Germering, 30th of September, 2020

### **ALLNET Medical Display with PoE**

#### **ALL\_MedTab\_10RK3288A80PoE**



The safety advice in the documentation accompanying the products shall be obeyed. The conformity to the above directive is indicated by the CE sign on the device.

The **ALL\_MedTab\_10RK3288A80PoE** conforms to the RED Directives of 2014/53/EU.

This equipment meets the following conformance standards:

#### **Safety:**

**EN62368-1:2014+A11:2017**

#### **Health:**

**EN62311:2008**

**EN50665:2017**

#### **EMC:**

**ETSI EN301 489-1 V2.2.3**

**EN55032:2015**

**ETSI EN301 489-17 V3.1.1**

**EN55035:2017**

#### **Radio:**

**ETSI EN300 328 V2.2.2**

**ETSI EN300 440 V2.2.1**

**ETSI EN301 893 V2.1.1**

#### **LVD:**

**EN60950-1:2006+A11:2009+A1:2010+A12:2011+A2:2013**

This equipment is intended to be operated in all countries.

This declaration is made by

ALLNET GmbH Computersysteme  
Maistraße 2  
82110 Germering  
Germany

Germering, 30.09.2020

  
\_\_\_\_\_  
Wolfgang Marcus Bauer  
CEO